

# VERIFICATION OF CALIBRATION REPORT DEC 07 2010

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130131✓</u>	
Name <u>Shayne Calt</u>		ID# <u>4559✓</u>	Date <u>11-4-10✓</u>
A	Agency <u>AST</u>	Phone # <u>733-2256</u>	
Instrument Location <u>Talkootna</u>			
Alco S/N <u>D119594✓</u>		Target Value <u>.078</u>	High Pressure <u>1000</u>
B	Alco Test Value Average <u>.074✓</u>	<u>.074✓</u>	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>			
(OVER)			

BMB  
12/16/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of DEC, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130131

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130131

NOVEMBER 04, 2010

OPERATOR'S NAME:  
CALT SHAWNE M  
OPERATOR'S NUMBER: 4559  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
U

D.L. #: U  
DEPT/AGENCY: TAL0  
CASE/REPORT: U  
TEST TYPE: U  
ALCO TARGET VALUE: .073  
ALCO S/N: D119594

--- BREATH ANALYSIS ---

.073 ADJUSTED FOR 28.65 in  
ALCO TARGET .074 13:36  
BLANK TEST .000 13:37  
INTERNAL STANDARD (VERIFIED) 13:37  
ALCO TV 28.65 in .074 13:37  
BLANK TEST .000 13:38  
SUBJECT SAMPLE .000 13:40  
BLANK TEST .000 13:40  
ALCO TV 28.65 in .074 13:41  
BLANK TEST .000 13:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130131

NOVEMBER 04, 2010  
TIME 13:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 49c  
BAROMETER: 28.65 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

"\*22'()\*+,-./0123456789:;<=>?@AB CDEFG  
HIJ KLMNOPQRSTU VWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxy z{ }~"